



walter insegno
genova
16122
Italy

137 Harley Street, London W1G 6BG
Tel: +44 20 7034 3070 Fax: +44 20 7034 3071
Email: fmfcertification@fetalmedicine.com

Feb 14, 2012

FMF ID: 17499

Measurement of uterine artery PI

Dear walter insegno,

Thank you for submitting your logbook of 3 images. I examined each image for the following criteria: sagittal section, angle of insonation angle less than 30 degrees, velocity more than 60 cm/s, and sample volume 2.0 mm (see table below).

<i>Criteria</i>	<i>Image number</i>		
	<i>1</i>	<i>2</i>	<i>3</i>
Sagittal section	✓	✓	✓
Insonation angle < 30°	✓	✓	✓
Velocity > 60 cm/s	✓	✓	✓
Sample Volume 2.0 mm	✓	✓	✓

Legend: ✓ = pass ✗ = fail -- = not possible to assess

Your images are satisfactory. Your software license will include the assessment of **uterine artery flow** in the risk calculation for preeclampsia and your name will appear on the list of sonographers certified for assessment of uterine artery flow.

Best wishes,

Olga Torres Khoury